



America

CERTIFICATE

No. U8 05 10 41603 006

Holder of Certificate: TissueLink Medical, Inc.

One Washington Center, Ste 400
Dover, NH 03820
USA

Certification Mark:



Product:

Medical Products
HF-surgery generator with rotary peristaltic pump

The product was voluntarily tested according to the relevant safety requirements and mentioned properties. It can be marked with the certification mark shown above. See also notes overleaf.

Test report no.:

052-098578-000



Date, 2005-10-18

Page 1 of 2



America

CERTIFICATE
No. U8 05 10 41603 006

Model(s): **AQUAMANTYS 40-402-1**

Parameters:

Rated Input Voltage:	115 Vac
Rated Frequency:	50/60 Hz
Rated Input Power:	420 VA
Rated Output Power:	200 W at 100 Ohm
Rated Load:	100 Ohm
Rated Output Frequency:	370 kHz
Protection Class:	I
Applied Part Type:	CF with defibrillation-proof

Remarks:

The certificate is valid for USA and Canada.
Other countries might have other requirements,
which were not part of this certification.

Tested according to: UL 60601-1:2003
CAN/CSA C22.2 No.601.1-M90

Production Facility(ies): 21988